

**Finn Chamber®**  
for epicutaneous testing

**Finn Chambers® on Scanpor®**  
for epicutaneous testing

**Finn Chambers® on Scanpor®**

**Finn Chamber®** is a patch test device which provides good occlusion because of the chamber design. Made of aluminium, the 8 mm inner diameter provides a 50 mm<sup>2</sup> area and about 20 microlitre volume. Larger chambers (12 mm inner diameter) and chambers with polypropylene coating (8, 12, 18 mm) are available for special purposes. Finn Chambers are mounted on Scanpor® tape from Actavis Norway AS, Norgesplaster Facility, Norway, with protective paper backing which is easily peeled away.

**Finn Chambers on Scanpor** are available in strips of 10 (2x5), 5 (1x5) and 1 chambers. The strips of 10 chambers are practical when testing with a large number of substances, e.g. with routine tests. Smaller strips are suitable for small test series and individual tests. The strips should not be cut between the chambers, as the remaining adhesive area may be too small.

For locating the test sites a special device, **Reading Plate**, is recommended. **Finn Chamber Tray** aids the handling of test tapes at the application of test substances.

Allergic reactions to aluminium and Scanpor tape are rare. However, occasional cases of contact sensitivity to aluminium e.g. due to vaccination or hyposensitization of hay fever patients with aluminium precipitated antigens have been reported.

The skin may react to the removal of the tests with a slight mechanical irritation, indicated by erythema on the area covered by the tape.

**Test substances** are usually applied in

petrolatum. The concentrations of allergens in most standard series are suitable for Finn Chambers. When using uncommon test substances the administering physician should choose carefully the substances and concentrations. It is advisable to use low concentrations with irritating test substances due to tight occlusion provided by the chamber.

Due to the incompatibility of aqueous mercuric solutions with aluminium, polypropylene coated Finn Chambers should be used when testing mercuric compounds. Aluminium may enhance the polymerization of acrylate monomers and false negative reactions have been noticed with acetone solutions of ethyl cyanoacrylate glue.

**Instructions for use**

Finn Chambers on Scanpor are intended for use by or under the supervision of a physician.

**Application of the test substances.** Mark identification on the top of each tape to show the order of the test substances throughout the testing procedure. Remove the protective paper and place the tape on the desk or the Finn Chamber Tray with the chambers up (Fig. 1). Keep a narrow strip of the protective paper on the tape until the tape has been attached onto the skin.

**Semisolids** (e.g. petrolatum as the vehicle) are applied directly into the chamber, filling more than half the chamber volume (a bar of about 5-6 mm if the diameter is 2 mm). Do not use filter paper discs with semisolids.

For **liquids** place a filter paper disc in the chamber. Moisten the disc thoroughly without surplus. Excess liquid should be removed e.g. with

porous paper. Place the test onto the skin within a few minutes. Do not let the filter paper disc dry, because this may result in weak or false negative reactions and the disc tends to slip out during application onto the back.

**Placing the tests onto the skin.** The patient shall stand or sit in a relaxed normal position, the back slightly bent forwards. The tests are applied to healthy skin on the back, which is to be free of ointments and excessive sebum. It is recommended that the patient takes a shower or bath in the morning before testing. If necessary, the skin can be cleaned e.g. by alcohol.

Apply the tape starting with the lower part and pressing the chambers from below to let the air escape. After having applied the tape this way, press each chamber containing a semisolid gently with the finger to get an even distribution of the test substance. Rub the tape gently but firmly with the palm against the skin, especially on the corners, to ensure good adherence. The patient should refrain from vigorous activities and taking showers during the first day after application of the tests.

The skin should be marked e.g. with nontoxic ink or tape.

**Removing the tests.** The tests are removed after one or two days' exposure. Whenever possible the patient should return to the clinic for removal of the tests. If the patient removes the tests himself, he should be advised to destroy the remnants properly.

**Check test sites** immediately after removal of the tests. A ringshaped depression around each test (Fig. 2) verifies the occlusion and validates the test, especially in the case of negative reactions.

**Reading the test reactions.** The tests should be read not less than 20 minutes after removal. The subsequent readings should be done 3-7 days after the application of the tests. Observing the course of the reactions helps in differentiating the allergic reactions from the irritant ones.

To locate and identify the reactions align the top left corner of the Reading Plate either with the ink mark on the skin or tape (Fig. 2). The Reading Plate serves as a template with the holes indicating the test sites.

**Storage** at room temperature, cold tape adheres less well.




Fig. 1




Fig. 2